

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2332600303
DATE : 03.02.2026
REQ No : 1032600496
REQ TITLE : PR for the Calibration of Bar
DELIVERY LOCATION
CLOSING DATE : 09.02.2026
CLOSING TIME : 10:00:00

SL NO	ITEM CODE	SERVICE ID	DESCRIPTION	QTY	UOM	BRAND/ORIGIN	WARRANTY	UNIT PRICE	TOTAL PRICE
1		ITEM-00001	CALIBRATION OF BAROMETRIC PRESSURE / TEMPERATURE & HUMIDITY DATA LOGGER INSTRUMENT DETAILS MAKE: EXTECH MODEL: SD 700 RANGE & SPECIFICATIONS BAROMETRIC PRESSURE: 10 TO 1100 HPA; 7.5 TO 825.0 MMHG; 0.29 TO 3248 INHG RESOLUTION: 0.1 HPA, 0.1 MMHG, 0.01 INHG TEMPERATURE: 32 TO 122°F (0 TO 50°C) RESOLUTION/BASIC ACCURACY: 0.1°F (°C) / ±1.5°F (0.8°C) RELATIVE HUMIDITY: 10 TO 90% RH RESOLUTION/BASIC ACCURACY: 0.1% / ± (4% OF RDG + 1% RH) # CALIBRATION CERTIFICATION SHALL BE FROM ISO/ IEC 17025 ACCREDITATION LAB WITH THE RELEVANT SCOPE. # UNCERTAINTY CALCULATION REQUIRED IN %# 2 YEARS CALIBRATION VALIDITY CONTACT PERSON : MAHRA ALI MOHAMMED ALMUTTAWA (0552227744)	1	EA				
2		ITEM-00002	CALIBRATION OF LOAD CYCLE TEST SYSTEM (HEAT CYCLE TEST) INSTRUMENT DETAILS MAKE: SA ASSOCIATES MODEL: LCTS 2000.30 SERIAL: 310321 MEASURE MENT RANGE TRANSFORMER CURRENT OUTPUT: 300 TO 2000 AMP TIMER: 0 TO 9999 (H : M : S)	1	EA				

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2332600303
DATE : 03.02.2026
REQ No : 1032600496
REQ TITLE : PR for the Calibration of Bar
DELIVERY LOCATION
CLOSING DATE : 09.02.2026
CLOSING TIME : 10:00:00

SL NO	ITEM CODE	SERVICE ID	DESCRIPTION	QTY	UOM	BRAND/ORIGIN	WARRANTY	UNIT PRICE	TOTAL PRICE
			S) TEMPERATURE MEASURING SENSORS: 0 TO 150°C (3 NOS.)# CALIBRATION CERTIFICATION SHALL BE FROM ISO/ IEC 17025 ACCREDITATION LAB WITH THE RELEVANT SCOPE.# UNCERTAINTY CALCULATION REQUIRED IN % FOR CURRENT & TEMPERATURE# ONSITE CALIBRATION REQUIRED# 2 YEARS CALIBRATION VALIDITY CONTACT PERSON : MAHRA ALI MOHAMMED ALMUTTAWA (0552227744)						
3		ITEM-00003	CALIBRATION OF HOT SET HEATING OVEN INSTRUMENT DETAILS# MAKE: MEMMERT# MODEL: UN55# SERIAL: B222.0847 RANGE# TEMPERATURE: 0 - 300°C# TIMER: 0 TO 9999# CALIBRATION CERTIFICATION SHALL BE FROM ISO/ IEC 17025 ACCREDITATION LAB WITH THE RELEVANT SCOPE.# UNCERTAINTY CALCULATION REQUIRED IN %# ONSITE CALIBRATION REQUIRED# 2 YEARS CALIBRATION VALIDITY CONTACT PERSON : MAHRA ALI MOHAMMED ALMUTTAWA (0552227744)	1	EA				

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2332600303
DATE : 03.02.2026
REQ No : 1032600496
REQ TITLE : PR for the Calibration of Bar
DELIVERY LOCATION
CLOSING DATE : 09.02.2026
CLOSING TIME : 10:00:00

SL NO	ITEM CODE	SERVICE ID	DESCRIPTION	QTY	UOM	BRAND/ORIGIN	WARRANTY	UNIT PRICE	TOTAL PRICE
			744)						

TOTAL AMOUNT IN WORDS:

TOTAL AMOUNT:

SPECIAL NOTES / INSTRUCTIONS:

PR for the Calibration of Barometric Pressure / Temp <(>&<>) Humidity

Data Logger, Load Cycle Test System, Hot Set Heating Oven

Contact Person : Ms. Mahra Ali Mohammed AlMutawa (0552227744)

STANDARD TERMS & CONDITIONS

- 1) Prices should be 'DDP' delivery duty paid at DEWA stores.
- 2) Quotation to be submitted only in local currency U.A.E Dirhams
- 3) DEWA Standard payment terms is '30 days credit' from the date of acceptance of material
- 4) No DEWA staff or his or her relatives up to third degree should have ownership or partnership in your company, and your participation in DEWA tenders / RFQs should not constitute a Conflict or perceived Conflict of Interest.
- 5) The offered product and/ or services in the Quotation, shall be conforming and in accordance with DEWA Energy Management Policy & EnMS Manual.
- 6) Confidentiality Clause: The Supplier/Service Provider shall treat all information and data (excluding open data) contained in the Purchase Order or obtained by the Supplier/Service Provider in connection with the execution of Works as proprietary and strictly confidential. In particular, the Supplier/ Service Provider shall not publish or disclose any data or information including but not

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER	: 2332600303
DATE	: 03.02.2026
REQ No	: 1032600496
REQ TITLE	: PR for the Calibration of Bar
DELIVERY LOCATION	
CLOSING DATE	: 09.02.2026
CLOSING TIME	: 10:00:00

limited to personally identifiable information, government data, project details, specifications, drawings or photographs concerning the Works to any third parties without the prior written consent of the Employer.

If the Supplier/Service Provider is obligated to obtain NOCs / Permits /Permissions for the Project from Government agencies as per the terms and scope of the Contract, the Supplier/Service Provider shall submit only the specific data required for the purpose to the authorized team of the agency, through their prescribed channel / system, subject to information security requirements

The Supplier/Service Provider shall comply with all the applicable data protection laws and regulations.

SUPPLIER'S REMARKS :

SUPPLIER'S SIGNATURE AND STAMP