

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492400336
DATE : 12.06.2024
REQ No : 1172400402
REQ TITLE : Laboratory Equipments for Tra
DELIVERY LOCATION : 0051 - CENTR STROE(TRN)
CLOSING DATE : 21.06.2024
CLOSING TIME : 10:00:00

SL NO	ITEM CODE	SERVICE ID	DESCRIPTION	QTY	UOM	BRAND/ORIGIN	WARRANTY	UNIT PRICE	TOTAL PRICE
1	7000000013		<p>TECHNICAL REQUIREMENTS FOR LABORATORY DISH WASHERUSAGE/APPLICATION: FORAUTOMATED WASHING AND DRYING LABORATORY GLASSWARECONSTRUCTION TYPE:FREESTANDING, WITHOUT LID, STAINLESS STEEL CASING,AUTO CLOSE DOOR, AUTOMATIC DOOR LOCK, BUZZER, ACOUSTIC SIGNAL AT END OFPROGR AMCAPACITY: 120- 180 LITERS OF WASH CABINET USABLE CAPACITY128 NUMBERS LABORATORY PHIALS /FLASKS PER CYCLE98 NUMBERS OF PIPETTES PER CYCLE130 NUMBERS OF MIXED LOAD CONSISTING OF LABORATORY PHIALS AND PIPETTESORMIXED LOAD CONSISTING OF LABORATORY GLASSWARE AND PHIA LSDRYING:INTEGRATED DISPENSER PUMPDRAWER FOR SUPPLY CANISTERS (3 X 5 L OR 2 X 10 L)ELECTRICAL CONNECTION 3N AC 400 V, 50 HZTOTAL RAT ED LOAD [KW] 9.3PERFORMANCE DATA: CIRCULATION PUMP, QMAX 500 L/MINSHORTEST PROGRAM RUNNING TIME IN MIN 20MAXIMUM</p>	1	EA				

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492400336
DATE : 12.06.2024
REQ No : 1172400402
REQ TITLE : Laboratory Equipments for Tra
DELIVERY LOCATION : 0051 - CENTR STROE(TRN)
CLOSING DATE : 21.06.2024
CLOSING TIME : 10:00:00

SL NO	ITEM CODE	SERVICE ID	DESCRIPTION	QTY	UOM	BRAND/ORIGIN	WARRANTY	UNIT PRICE	TOTAL PRICE
			<p>FINAL RINSE TEMPERATURE IN °C 93 CONTROL SYSTEM: TOUCH CONTROL PROGRAM SELECTION: TOUCH ON METAL TOUCH TYPE DIRECT-ACCESS PUSHBUTTONS MAX. DELAY START : 24 H PROGRAMMABILITY: WASH CYCLES SHALL BE PROGRAMMABLE STANDARD ELECTRICAL CONNECTION: ELECTRICAL CONNECTION AC 230V 50HZ ELECTRICAL CONNECTION 3N AC 400V 50HZ TOTAL RATED LOAD [KW] 9.3 WATER CONNECTION/DRAINAGE: COLD WATER CONNECTION [1 NUMBER] HOT WATER CONNECTION [NUMBER] DRAIN PUMP PROTECTION AGAINST FLOODING: IN THE EVENT OF PROBLEMS, THE SYSTEM STOP WATER INTAKE AUTOMATICALLY. DUTY CYCLE: 100% OPERATING TEMPERATURE: 5 TO 40 DEG HUMIDITY RANGE: 20 TO 85% RH WASH CABINET, USABLE HEIGHT: 520 MM APPROX WASH CABINET, USABLE WIDTH: 530 MM APPROX. WASH CABINET, DEPTH OF UPPER BASKET: 474 MM APPROX WASH CABINET, DEPTH</p>						

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492400336
DATE : 12.06.2024
REQ No : 1172400402
REQ TITLE : Laboratory Equipments for Tra
DELIVERY LOCATION : 0051 - CENTR STROE(TRN)
CLOSING DATE : 21.06.2024
CLOSING TIME : 10:00:00

SL NO	ITEM CODE	SERVICE ID	DESCRIPTION	QTY	UOM	BRAND/ORIGIN	WARRANTY	UNIT PRICE	TOTAL PRICE
			<p>OF LOWER BASKET 520 MM APPROX.NET WEIGHT: 100 KG APPROX.ADDITIONAL REQUIREMENTS: STAINLESS-STEEL LID,UPPER BASKET FOR LOADING OF VARIOUS MODULES AND INSERTSLOWER BASKET FOR LOADING OF VARIOUS MODULES AND INSERTSMODULE 3X6, FOR LOADING OF 18 QTY NARROW NECKED GLASSWARE AND MEASURINGCYLINDERSMODULE 2X4, FOR LOADING OF 8 NARROW NECKED GLASSWARES AND MEASURINGCYLINDERSOPEN BASKET FOR BEAKERS AND WIDE NECKED GLASSWARESINSERT ¼ FOR APPROX 200 TEST TUBES, MAX 12X105MMLIQUID DETERGENT # 5 LTR CANISTERLIQUID NEUTRALIZER # 5 LTR C ANISTERNOZZLE INJECTOR MODULES AND LOAD CARRIERS FOR PIPETTES, PHIALS ANDMEASURING CYLINDERSDETERGENT LEVEL SENSORUNIVERSAL SUPPORT FOR INJECTION SYSTEMSPIPETTE ADAPTER FOR MULTI-NOZZLE SYSTEMSACCEPTABLE MODELS:</p>						

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492400336
DATE : 12.06.2024
REQ No : 1172400402
REQ TITLE : Laboratory Equipments for Tra
DELIVERY LOCATION : 0051 - CENTR STROE(TRN)
CLOSING DATE : 21.06.2024
CLOSING TIME : 10:00:00

SL NO	ITEM CODE	SERVICE ID	DESCRIPTION	QTY	UOM	BRAND/ORIGIN	WARRANTY	UNIT PRICE	TOTAL PRICE
			MIELE PG 8583 CD WITH ACCESSORIES / FROM JTHE MIELE COMPANY UK OR GW6290DS OR EQUIVALENT DELIVERY LOCATION: OIL TEST LABORATORY AT WARSAN , DUBAI INSTALLATION: INSTALLATION & NECESSARY PLUMBING WORK SHALL BE INCLUDED IN OFFER. TRAINING: COMPREHENSIVE TRAINING FOR OPERATION SHALL BE PROVIDED BY CERTIFIED ENGINEER CONSUMABLES: CONSUMABLES NECESSARY FOR ONE YEAR OF OPERATION SHALL BE SUPPLIED WITH DISHWASHER WARRANTY: 1 YEAR INCLUDING SPARE PARTS & ENGINEERING FEES						
2	7000000013		TECHNICAL REQUIREMENTS FOR LABORATORY AIR COMPRESSOR USAGE/APPLICATION: LABORATORY USE AIR FLOW RATE: 20 TO 135L/MIN WITH A MAXIMUM PRESSURE OF 8 BAR (116 PSI) VOLTAGE: 230 VAC FREQUENCY : 50 HZ MOTOR : 0.33 TO 1 HP RECEIVER VOLUME: 8 LITERS NOISE LEVEL : <65	1	EA				

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492400336
DATE : 12.06.2024
REQ No : 1172400402
REQ TITLE : Laboratory Equipments for Tra
DELIVERY LOCATION : 0051 - CENTR STROE(TRN)
CLOSING DATE : 21.06.2024
CLOSING TIME : 10:00:00

SL NO	ITEM CODE	SERVICE ID	DESCRIPTION	QTY	UOM	BRAND/ORIGIN	WARRANTY	UNIT PRICE	TOTAL PRICE
			DB(A)/1MAIR COMPRESSOR SHALL BE FITTED WITH STANDARD WITH A 8 LITER CAPACITYPRESSURE VESSEL TANKCOMPRESSOR SHALL BE 100 % OIL FREEDUTY CYCLE: 10 0%ACCESSORIES INCLUDING AUTOMATIC DRAIN WITH A DRAIN-PIPE, SINTER FILTERAND 1 LT DRAIN BOTTLE SHALL BE SUPPLIED WITH COMPRESSOR.TO B E FITTED WITH PRESSURE REGULATOR ON TANK TO CONTROL OUTLET PRESSURETO 5 BAROPERATING TEMPERATURE: 5 TO 40 DEG CHUMIDITY RANGE: 20 TO 85% RHACCEPTABLE MODELS: SYNERGY MAKE SYN-8L OR EQUIVALENT						
3	7000000013		TECHNICAL REQUIREMENTS FOR DIGITAL CONTACT THERMOMETERUSAGE/APPLICATION: LABORATORY APPLICATIONTHERMOMETER TYPE: RESISTIVE CONTACT OR RESISTIVE THERMOMETER DEVICE(RTD) PLATINUM TRACEABLE THERMOMETERTRACEABILITY STANDARDS:ASTM E2877 CLASS A, IEC 751, ITS 90 , AST	1	EA				

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492400336
DATE : 12.06.2024
REQ No : 1172400402
REQ TITLE : Laboratory Equipments for Tra
DELIVERY LOCATION : 0051 - CENTR STROE(TRN)
CLOSING DATE : 21.06.2024
CLOSING TIME : 10:00:00

SL NO	ITEM CODE	SERVICE ID	DESCRIPTION	QTY	UOM	BRAND/ORIGIN	WARRANTY	UNIT PRICE	TOTAL PRICE
			<p>M E644,ASTM E1137 ,ASTM , ASTM D8278, ITS 90DISPLAY RESOLUTION: 0,001 °CRANGE -40°C TO 140°CACCURACY 0.02 °CREADING °CDISPLA Y TYPE: 2 LINE LIQUID CRYSTAL DISPLAY + BACKLIGHTPROBE: 215MM LENGTH (APPROX.), MATERIAL 304 STAINLESS STEELDISPLAY ACCURACY : ±20 MK (±0,02 °C) FOR COMBINED PROBE AND SENSORSSENSOR TYPE: RTD, SUCH AS A PRT OR THERMISTORDRIFT: LESS THAN 10 MK (0,01 °C) PER YEARRESPONSE TIME: LESS THAN OR EQUAL TO 3 S, AS DEFINED IN SPECIFICATIONASTM E1137/E1137MLINEARITY: (±0,01 °C) 10 MK OVER RANGE OF INTENDED USEINTERFACE USBPOWER 9V, MAINS ADAPTORACCESSORY ADAPTER TO INSERT AN TT3B THERMOMETER IN A 12.5 MM THER MOMETEROPENING.CALIBRATION REPORT:CALIBRATION REPORT WHICH SHOULD BE TRACEABLE TO ANATIONAL CALIBRATION OR METROLOGY STANDARDS BODY ISSUED BY AN ISO/IEC17025 ACCREDITED CALIBRATION LABORATORY WITH DEMONSTRATED</p>						

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492400336
DATE : 12.06.2024
REQ No : 1172400402
REQ TITLE : Laboratory Equipments for Tra
DELIVERY LOCATION : 0051 - CENTR STROE(TRN)
CLOSING DATE : 21.06.2024
CLOSING TIME : 10:00:00

SL NO	ITEM CODE	SERVICE ID	DESCRIPTION	QTY	UOM	BRAND/ORIGIN	WARRANTY	UNIT PRICE	TOTAL PRICE
			<p>COMPETENCY INTEMPERATURE CALIBRATION SHALL BE SUPPLIED WITH THERMOMETER. 17025ACCREDITED CALIBRATION LABORATORY WITH TEMPERATURE CALIBRATION IN ITSACCREDITATION SCOPE AND WHICH MEETS THE STATED MEASUREMENT UNCERTAINTYWOULD MEET THIS REQUIREMENT.THE CALIBRATION DATA SHALL INCLUDE AT LEAST 3 CALIBRATION TEMPERATURESINC LUDING 0 °C AND 2 OTHER POINTS INCLUDING THE TEST TEMPERATURE OF USEAND STATE THE IMMERSION DEPTH UNDER WHICH THIS WAS CALIBRATED AN D THERESULTING UNCERTAINTY.ACCEPTABLE MODEL/ MAKE TT3B FROM TAMSON INSTRUMENTS OR EQUIVALENT</p>						

TOTAL AMOUNT IN WORDS:

TOTAL AMOUNT:

SPECIAL NOTES / INSTRUCTIONS:

MATERIAL AND OR SERVICE PROVIDED TO DEWA SHOULD PREFERABLY BE ENERGY EFFICIENT AND ENVIRONMENT FRIENDLY.

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492400336
DATE : 12.06.2024
REQ No : 1172400402
REQ TITLE : Laboratory Equipments for Tra
DELIVERY LOCATION : 0051 - CENTR STROE(TRN)
CLOSING DATE : 21.06.2024
CLOSING TIME : 10:00:00

Laboratory Equipments for Transformer Oil Test Lab.

STANDARD TERMS & CONDITIONS

- 1) Prices should be 'DDP' delivery duty paid at DEWA stores.
- 2) Quotation to be submitted only in local currency U.A.E Dirhams
- 3) DEWA Standard payment terms is '30 days credit' from the date of acceptance of material
- 4) No DEWA staff or his or her relatives up to third degree should have ownership or partnership in your company, and your participation in DEWA tenders / RFQs should not constitute a Conflict or perceived Conflict of Interest.
- 5) The offered product and/ or services in the Quotation, shall be conforming and in accordance with DEWA Energy Management Policy & EnMS Manual.

SUPPLIER'S REMARKS :

SUPPLIER'S SIGNATURE AND STAMP