

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492500766
DATE : 18.11.2025
REQ No : 1172501235
REQ TITLE : True-RMS Clamp Meter
DELIVERY LOCATION : 0051 - CS WRSN, TRNSMSN
CLOSING DATE : 24.11.2025
CLOSING TIME : 10:00:00

SL NO	ITEM CODE	SERVICE ID	DESCRIPTION	QTY	UOM	BRAND/ORIGIN	WARRANTY	UNIT PRICE	TOTAL PRICE
1	7000000013		LINE ITEM-1:TRUE-RMS CLAMP METER DIGITAL WITH FLEXIBLE CURRENT PROBESPECIFICATION:VOLTAGE: RANGE 1000 V AC / DC; RESOLUTION 0.1 V (600 V),1 V (1000 V);ACCURACY 1% ±5 DIGITS (20 HZ TO 500 HZ)CURRENT VIA JAW: RANGE 999.9 A AC / DC, RESOLUTION 0.1 A ACCURACY ACCURREN TS 2% ±5 DIGITS; ACCURACY DC CURRENTS 2% ±5 DIGITS (10-100HZ) &2.5% ±5 DIGITS (100-500HZ)ACCURACY AC CURRENTS 3%CURRENT VIA 18-INCH FLEXIBLE CURRENT PROBERANGE 2500 A AC; RESOLUTION 0.1 A (600A) &1A(2500A); ACCURACY ACCURREN TS 3% ±5 DIGITS (5-500HZ)FLEXIBLE CURREN T PROBE CABLE LENGTH (HEAD TO ELECTRONICS CONNECTOR) 1.8MTHE FLEXIBLE PROBE: EXPANDS THE MEASUREMENT RANGE TO 2500 A AC ALLOWSFREQUE NCY MEASUREMENT TO 500 HZ WITH JAW AND FLEXIBLE PROBERESISTANCE: 60 K# WITH CONTINUITY DETECTION; ACCURACY 1% ±5 DIGITSDC MV: 500 MV	2	EA				

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492500766
DATE : 18.11.2025
REQ No : 1172501235
REQ TITLE : True-RMS Clamp Meter
DELIVERY LOCATION : 0051 - CS WRSN, TRNSMSN
CLOSING DATE : 24.11.2025
CLOSING TIME : 10:00:00

SL NO	ITEM CODE	SERVICE ID	DESCRIPTION	QTY	UOM	BRAND/ORIGIN	WARRANTY	UNIT PRICE	TOTAL PRICE
			MEASUREMENT RANGE TO INTERFACE WITH OTHER ACCESSORIESCAPACITANCE MEASUREMENT: 1000 #F; ACCURACY 1% ±4 DIGITSFREQUENCY VIA JAW & CUR RENT PROBE: 5.0 HZ TO 500.0 HZ, RESOLUTION 0.1HZ, ACCURACY 0.5% ±5 DIGITSMAXIMUM VOLTAGE BETWEEN ANY TERMINAL AND EARTH GROUND 1000 VOPERATING TEMPERATURE#0-50CKEY FEATURES# MEASURES AC/DC CURRENT WITH INCLUDED FLEXIBLE CURRENT PROBE# VFD LOW PASS FILTER FOR ACCUR ATE MEASUREMENTS OF NON-LINEAR SIGNALS# REMOTELY LOG, TREND AND MONITOR TO PINPOINT INTERMITTENT FAULTS# INRUSH MEASUREMENT TO FILTE R OUT NOISE AND CAPTURE STARTING CURRENTACCESSORIES:MAGNETIC HANGING STRAP, TEST LEADS, BATTERIES (INSTALLED), SOFT CARRYINGCASECOAT ED INSTRUCTION CARD AND SAFETY INFORMATION SHEETCALIBRATION CERTIFICATE AND WARRANTY VALID FOR 2 YEARSSAFETY						

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492500766
DATE : 18.11.2025
REQ No : 1172501235
REQ TITLE : True-RMS Clamp Meter
DELIVERY LOCATION : 0051 - CS WRSN, TRNSMSN
CLOSING DATE : 24.11.2025
CLOSING TIME : 10:00:00

SL NO	ITEM CODE	SERVICE ID	DESCRIPTION	QTY	UOM	BRAND/ORIGIN	WARRANTY	UNIT PRICE	TOTAL PRICE
			RATING: CLAMP CARRIES A CAT III 1000 V, CAT IV 600 V. COMPLY WITH IEC AND APPLICABLE INTERNATIONAL STANDARDS FOR SAFETY, IP RATING, ELECTROMAGNETIC COMPATIBILITY (EMC), RADIO FREQUENCY CERTIFICATION ETC. NOTE: BIDDERS TO SHOW OFFERED SAMPLES AT DEWA WARSAN TEMC DURING THE RFX STAGE TO ENSURE COMPLIANCE WITH SPECIFICATIONS, QUALITY STANDARDS AND FUNCTIONAL CHECK.						
2	7000000013		LINE ITEM-2: AC / DC CURRENT CLAMP METER DIGITAL SPECIFICATION: MEASURING RANGE#0-40 A DC OR AC PK AUTOMATIC RANGE#4A/40A RESOLUTION#1mA IN 4 A RANGE, 10mA IN 40 A RANGE BASIC ACCURACY# $\pm 1.3\% + 5$ DIGIT TEMPERATURE COEFFICIENT # $\pm 0.05\%$ OF RDG / CFREQUENCY RANGE#40 HZ TO 400 HZ IN A COVER LOAD CAPACITY#150A SAFETY RATING#CAT III, 300V OPERATING TEMPERATURE#0-50C KEY FEATURES:# EASY-TO-MEASURE RANGE SWITCHING F	2	EA				

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492500766
DATE : 18.11.2025
REQ No : 1172501235
REQ TITLE : True-RMS Clamp Meter
DELIVERY LOCATION : 0051 - CS WRSN, TRNSMSN
CLOSING DATE : 24.11.2025
CLOSING TIME : 10:00:00

SL NO	ITEM CODE	SERVICE ID	DESCRIPTION	QTY	UOM	BRAND/ORIGIN	WARRANTY	UNIT PRICE	TOTAL PRICE
			OR AC/DC LEAKAGE CURRENT MEASUREMENTS FROM 10 MA TO 40 A# UP TO 1 MA CURRENT RESOLUTION FOR FAST CIRCUIT COMPARISON OF SENSORS AND CONTROL CIRCUITS# JAW DESIGN SUITABLE FOR HIGH ACCURACY, SMALL DC CURRENT MEASUREMENTS TO ENSURE INTERFERENCE FROM OTHER CIRCUITS# DC AUTO-ZERO BUTTON ENSURES CORRECT MEASUREMENT EVERY TIME# HAS AUTOMATIC POWER-OFF FUNCTION FOR SAVING BATTERY POWER# AMMETER IS SMALL IN SIZE, LIGHT IN WEIGHT AND EASY TO CARRY# EASY DISPLAY HOLD FEATURE MAKES IT EASY TO CAPTURE MEASUREMENTS AT INACCESSIBLE LOCATIONS# ACCESSORIES: # MAGNETIC HANGING STRAP, TEST LEADS, BATTERIES (INSTALLED), SOFT CARRYING CASE# COATED INSTRUCTION CARD AND SAFETY INFORMATION SHEET# CALIBRATION CERTIFICATE AND WARRANTY VALID FOR 2 YEARS COMPLY WITH IEC AND APPLICABLE INTERNATIONAL STANDARDS						

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492500766
DATE : 18.11.2025
REQ No : 1172501235
REQ TITLE : True-RMS Clamp Meter
DELIVERY LOCATION : 0051 - CS WRSN, TRNSMSN
CLOSING DATE : 24.11.2025
CLOSING TIME : 10:00:00

SL NO	ITEM CODE	SERVICE ID	DESCRIPTION	QTY	UOM	BRAND/ORIGIN	WARRANTY	UNIT PRICE	TOTAL PRICE
			FOR SAFETY , IP RATING, ELECTROMAGNETIC COMPATIBILITY (EMC), RADIO FREQUENCY CERTIFICATION ETC. NOTE: BIDDERS TO SHOW OFFERED SAMPLES AT DEWA WARS AN TEMC DURING THE RFX STAGE TO ENSURE COMPLIANCE WITH SPECIFICATIONS, QUALITY STANDARDS AND FUNCTIONAL CHECK.						
3	7000000013		LINE ITEM-3: AC / DC MINI CLAMP METER SPECIFICATION: MEASURING RANGE#0.0-1.0 A AND 0~10A RESOLUTION#1mA BASIC ACCURACY# $\pm 1.3\%$ +5 DIGITS FREQUENCY RANGE#40 HZ TO 400 HZ IN AC OVERLOAD CAPACITY#150A SAFETY RATING#C AT III, 300V OPERATING TEMPERATURE#0-50C# MULTIMETER CONNECTION FOR MEASURING READING DISPLAY# JAW DESIGN SUITABLE FOR HIGH ACCURACY, SMALL CURRENT MEASUREMENTS# ZERO BUTTON RESET ENSURES CORRECT MEASUREMENT EVERY TIME# HAS AUTOMATIC POWER-OFF FUNCTION FOR SAVING BATTERY POWER# TEST LEADS WITH BANANA PLUGS#	1	EA				

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492500766
DATE : 18.11.2025
REQ No : 1172501235
REQ TITLE : True-RMS Clamp Meter
DELIVERY LOCATION : 0051 - CS WRSN, TRNSMSN
CLOSING DATE : 24.11.2025
CLOSING TIME : 10:00:00

SL NO	ITEM CODE	SERVICE ID	DESCRIPTION	QTY	UOM	BRAND/ORIGIN	WARRANTY	UNIT PRICE	TOTAL PRICE
			CLAMP METER SMAL L IN SIZE, LIGHT IN WEIGHT AND EASY TO CARRY# EASY DISPLAY HOLD FEATUREACCESSORIES:# MAGNETIC HANGING STRAP, TEST LEADS, BATTERIES (INSTALLED), SOFTCARRYING CASE# COATED INSTRUCTION CARD AND SAFETY INFORMATION SHEET# CALIBRATION CERTIFICATE AND WARRANTY VALID FOR 2 YEARS COMPLY WITH IEC AND APPLICABLE INTERNATIONAL STANDARDS FOR SAFETY, IPRATING, ELECTROMAGNETIC COMPATIBILITY (EMC), RADIO FREQ UENCYCERTIFICATION ETC.NOTE: BIDDERS TO SHOW OFFERED SAMPLES AT DEWA WARSAN TEMC DURING THE RFXSTAGE TO ENSURE COMPLIANCE WITH SPECI FICATIONS, QUALITY STANDARDS ANDFUNCTIONAL CHECK.						

TOTAL AMOUNT IN WORDS:

TOTAL AMOUNT:

SPECIAL NOTES / INSTRUCTIONS:

True-RMS Clamp Meter

MATERIAL AND OR SERVICE PROVIDED TO DEWA SHOULD PREFERABLY BE ENERGY EFFICIENT AND ENVIRONMENT FRIENDLY.

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER	: 2492500766
DATE	: 18.11.2025
REQ No	: 1172501235
REQ TITLE	: True-RMS Clamp Meter
DELIVERY LOCATION	: 0051 - CS WRSN, TRNSMSN
CLOSING DATE	: 24.11.2025
CLOSING TIME	: 10:00:00

STANDARD TERMS & CONDITIONS

- 1) Prices should be 'DDP' delivery duty paid at DEWA stores.
 - 2) Quotation to be submitted only in local currency U.A.E Dirhams
 - 3) DEWA Standard payment terms is '30 days credit' from the date of acceptance of material
 - 4) No DEWA staff or his or her relatives up to third degree should have ownership or partnership in your company, and your participation in DEWA tenders / RFQs should not constitute a Conflict or perceived Conflict of Interest.
 - 5) The offered product and/ or services in the Quotation, shall be conforming and in accordance with DEWA Energy Management Policy & EnMS Manual.
 - 6) Confidentiality Clause: The Supplier/Service Provider shall treat all information and data (excluding open data) contained in the Purchase Order or obtained by the Supplier/Service Provider in connection with the execution of Works as proprietary and strictly confidential. In particular, the Supplier/ Service Provider shall not publish or disclose any data or information including but not limited to personally identifiable information, government data, project details, specifications, drawings or photographs concerning the Works to any third parties without the prior written consent of the Employer.
- If the Supplier/Service Provider is obligated to obtain NOCs / Permits /Permissions for the Project from Government agencies as per the terms and scope of the Contract, the Supplier/Service Provider shall submit only the specific data required for the purpose to the authorized team of the agency, through their prescribed channel / system, subject to information security requirements
- The Supplier/Service Provider shall comply with all the applicable data protection laws and regulations.

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492500766
DATE : 18.11.2025
REQ No : 1172501235
REQ TITLE : True-RMS Clamp Meter
DELIVERY LOCATION : 0051 - CS WRSN, TRNSMSN
CLOSING DATE : 24.11.2025
CLOSING TIME : 10:00:00

SUPPLIER'S REMARKS :

SUPPLIER'S SIGNATURE AND STAMP